

S/N: 001651

No.: **ICR Polska/VC/H210513**

Name and address of Applicant

Hangzhou Ruicheng Instrument Co., Ltd.
Building E, No.8 of Houshan Road, Yuhang District, Hangzhou City, Zhejiang Province, China

Name and address of manufacturer:

Hangzhou Ruicheng Instrument Co., Ltd.
Building E, No.8 of Houshan Road, Yuhang District, Hangzhou City, Zhejiang Province, China

Product name:

Thermo Shaker

Product types:

TS100, TCS10, TS200, TS300, TS400, TPS1000, TPS2000, TPS2000-H, TPS2000-L, TPS3000, TOS20, TOS20-C, TOS30, TOS30-C

Product trademark:

n/a

This document confirms that the product sample meets the requirements of the following standards:

- Related with Directives 2014/35/EU and 2014/30/EU:
EN 61326-1:2013
EN 61010-1:2010+A1:2019
EN 61010-2-051:2015

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

Test reports:

BSTXD2105212266602SR
BSTXD2105212266602ER

Tests conducted by:

BST Testing (Shenzhen)Co., Ltd.

Issue date:

31.05.2021

Expiration date:

30.05.2026

Remarks:

Document refers to the above mentioned product and its conformity in regards of above mentioned standard(s) was proven on test sample
Document was issued on voluntary basis and does not imply meeting all essential requirements, assessment of the series-production or any other restricted Notified Bodies conformity assessment procedure appropriate for the product
Document status can be checked: <https://cert.icrpolska.com/>

CE marking remarks:

CE mark is not sanctioned by the following document
CE mark given here as reference, can be only use by the manufacturer after applying all essential requirements from relevant directives



A handwritten signature in blue ink, appearing to be 'Rafał Kalinowski'.

Director: Rafał Kalinowski
Warsaw, 31.05.2021

